

# VERIFICATION OF CALIBRATION REPORT

NOV 12 2009

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130133</u> ✓
A	Name <u>BRYAN BARLOW</u>	ID# <u>4548</u> Date <u>11/12/09</u> ✓
	Agency <u>Alaska State Troopers</u>	Phone # <u>907-783-0972</u>
	Instrument Location <u>Girdwood AST</u> ✓	
B	Alco S/N <u>75260</u> ✓	Target Value <u>.079</u> ✓ High Pressure <u>600</u>
	Alco Test Value Average <u>.077</u> ✓	<u>.077</u> ✓
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
	Signature <u>[Signature]</u> ✓	

CAB  
12/1/09

(OVER)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.  
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.  
(3) I am the Scientific Director of the State Breath Alcohol Program.  
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.  
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.  
(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 21<sup>st</sup> day of Dec, 2009.

[Signature]  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130133

NOVEMBER 12, 2009

OPERATOR'S NAME: BARLOW BRYAN W  
OPERATOR'S NUMBER: 4548  
SUBJECT'S LAST NAME:  
OF CALIBRATION  
SUBJECT'S FIRST NAME/MI :  
VERIFICATION  
O.L. #: 1234567  
DEPT/AGENCY: GDMO  
CASE/REPORT: 09-000000  
TEST TYPE: V  
ALCO TARGET VALUE: .079  
ALCO S/N: 75260

### --- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.72 in  
ALCO TARGET .078 11:15  
BLANK TEST .000 11:17  
INTERNAL STANDARD VERIFIED 11:17  
ALCO TU 29.72 in .077 11:17  
BLANK TEST .000 11:18  
SUBJECT SAMPLE .000 11:19  
BLANK TEST .000 11:19  
ALCO TU 29.72 in .077 11:20  
BLANK TEST .000 11:21

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130133

NOVEMBER 12, 2009  
TIME 11:21

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 41c  
BAROMETER: 29.72 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKL MNOPQRSTU VWXYZ[\]^\_`'abcde fghijklmno  
pqrstu vwxyz{|}~■